



PROJECT ASSURE DIAMOND VERIFICATION INSTRUMENT STANDARD **TEST RESULTS**

Assessment Report for: Unimec SA _ SSEF **ASDI-500**



Prepared For: Luc Auer

Natural Diamond Council Belgium VOF

Hoveniersstraat 22 2018 Antwerpen

Received Date: May 8th, 2024

Assessment Dates: May 13th to May 15th, 2024

Testing ID Number: 2024-04 Report Date: May 28th, 2024

Approved by: Hamarand

Quinten Van Avondt

Lab Manager



Unimec SA / SSEF ASDI-500

Date:

May 28th, 2024

Testing ID:

2024-04

DIAMOND VERIFICATION INSTRUMENT

Manufacturer's Name: Unimec SA Instrument Model: ASDI-500 Serial Number: 6511000

Software Version:PLCv1.14 – HMIv1.10Lab Manager:Quinten Van AvondtAnalyst/Operator:Cindy De Plukker

Manufacturer stated diamond verification instrument description and features:

- Automatic stone feed
- Automatic stone classification
- Automatic stone sorting

Manufacturer stated diamond verification instrument limitations:

- Loose stones
- Round brilliant stone shape
- Stone size from 0.50 mm to 3.80 mm
- D-J colour
- Difference in stone diameter within a test parcel may not exceed +/- 0.05mm

INSTRUMENT PERFORMANCE ASSESSMENT

ASSESSMENT CRITERIA

The ASSURE testing methodology and performance metrics are dependent on the operational capabilities of the diamond verification instrument being tested. These are defined by the following three categories:

Category 1- Screen diamonds from synthetic diamonds. This category of device is intended for discrimination of diamonds from synthetic diamonds. It cannot distinguish diamonds from diamond simulants and therefore requires stones to be pre-screened to ensure no simulants are introduced into the device.

Category 2 – Screen diamonds from synthetic diamonds and diamond simulants. This category of device is intended for discrimination of diamonds from synthetic diamonds <u>and</u> diamond simulants. This device <u>cannot</u> distinguish synthetic diamonds from diamond simulants.

Category 3 – Screen diamond from synthetic diamonds from diamond simulants. This category of device is intended for discrimination of diamonds, synthetic diamonds and diamond simulants from each other. This device <u>can</u> distinguish synthetic diamonds from diamond simulants.

Instrument performance for classifying the different kinds of stones was assessed against:



	Diamond Verification Instrument Standard Part 1 – Diamond Verification Instrument for Screening Diamonds from Synthetic Diamonds (18 03 2024)
	Diamond Verification Instrument Standard Part 2 – Diamond Verification Instrument for Screening Diamonds from Synthetic Diamonds and Diamond Simulants (18 03 2024)
\boxtimes	Diamond Verification Instrument Standard Part 3 – Diamond Verification Instrument for Screening Diamonds, Synthetic Diamonds, and Diamond Simulants (18 03 2024)
	renced in sections 7.3 and 7.4 of the Diamond Verification Instrument Standard – General Requirements for Evaluation and Verification Instruments (18 03 2024). Any deviations from the Standard are noted below:
Notes	: The ASDI-500 has an upper size limit of 3.8 mm . Stones with greater diameter than 3.8 mm (30 out of 500) were
remov	red from the PRIMARY sample set prior to conducting performance testing.



Unimec SA / SSEF ASDI-500

Date:

May 28th, 2024

Testing ID:

2024-04

DEFINITIONS:

Diamond Accuracy	Diamond test stones correctly classified as Diamond.
Synthetic Diamond Accuracy	Synthetic Diamond test stones correctly classified as Synthetic Diamond.
Diamond Simulant Accuracy	Diamond Simulant test stones correctly classified as Diamond Simulant.
Diamond Referral Rate	Diamond test stones classified as Referral.
Synthetic Diamond Referral Rate	Synthetic diamond test stones classified as Referral
Simulant Referral Rate	Diamond simulant test stones classified as Referral
Diamond False Positive Rate	Synthetic diamonds /diamond simulants test stones incorrectly classified as Diamond.
Synthetic Diamond False Positive Rate	Diamonds/ Diamond simulants test stones incorrectly classified as Synthetic Diamond.
Diamond Simulant False Positive Rate	Diamonds /Synthetic Diamonds test stones incorrectly classified as Diamond Simulants.
Diamond False Negative Rate	Diamonds test stones incorrectly classified as Synthetic Diamonds or Diamond simulants.
Synthetic Diamond False Negative Rate	Synthetic Diamonds test stones incorrectly classified as Diamonds or Diamond simulants
Diamond Simulant False Negative Rate	Diamond Simulant test stones incorrectly classified as Diamonds or Synthetic Diamonds
Testing Speed	The average speed at which the diamond verification instrument evaluates the stones in the PRIMARY loose sample set , including set-up time (if any)
Operating Speed	For auto-loading diamond verification instruments only, the average speed at which stones are evaluated once the instrument achieves a steady-state. does not include set-up time.

TEST STONE SETS USED FOR EVALUATION

Loose, Polished Stone Test Sets	Diamond	Synthetic Diamond	Diamond Simulant
Primary Sample Set (>2.0 mm, D-J color)	\boxtimes	\boxtimes	\boxtimes
Supplementary Smalls Sample Set (1.0-2.0 mm, D-J color)	\boxtimes	\boxtimes	\boxtimes
Supplementary Ultra Smalls Sample Set (0.5mm > , < 1.0mm)	\boxtimes	\boxtimes	\boxtimes
Mounted, Polished Stone Test Sets	Diamond	Synthetic Diamond	Diamond Simulant
Primary Sample Set (>2.0 mm, D-J color)			
Supplementary Smalls Sample Set (1.0-2.0 mm, D-J color)			



Unimec SA / SSEF ASDI-500

Date:

May 28th, 2024

Testing ID:

2024-04

Notes: In Primary loose sample set, stones greater than 3.8 mm diameter were excluded from testing due to upper size limit for this instrument. This instrument cannot test mounted jewelry.

CLEANING PROCEDURE OF STONES PRIOR TO TESTING

Test stones sets are cleaned in an ultrasonic bath of isopropanol for 2 minutes and dried prior to testing to reduce grease and electrostatic charge, as per Section 8 of ASSURE Standard.

LABORATORY CONDITIONS AT TIME OF ASSESSMENT

Condition	Requirement	Actual
Temperature (°C)	18 to 25°C	24 °C
Humidity (%)	50 to 65%	54 %

RESULTS OF INSTRUMENT PERFORMANCE ASSESSMENT – LOOSE STONES

Performance Metric	Primary ^[1]	Uncertainty ^[2]	Smalls	Uncertainty ^[2]	Ultra-Smalls [4]
Diamond accuracy (%)	87.2	1.3	85.0	1.8	84.0
Synthetic diamond accuracy (%)	0.0 ^[3]	0.0	0.0[3]	0.0	0.0
Diamond simulant accuracy (%)	100.0	0.0	100.0	0.0	100.0
Diamond referral rate (%)	12.8	1.3	15.0	1.8	15.7
Synthetic diamond referral rate (%)	100.0 ^[3]	0.0	100.0 ^[3]	0.0	100.0
Diamond simulant referral rate (%)	0.0	0.0	0.0	0.0	0.0
Diamond false positive rate (%)	0.0	0.0	0.0	0.0	0.0
Synthetic diamond false positive rate (%)	0.0	0.0	0.0	0.0	0.0
Diamond simulant false positive rate (%)	0.0	0.0	0.0	0.0	0.2
Diamond false negative rate (%)	0.0	0.0	0.0	0.0	0.3
Synthetic diamond false negative rate (%)	0.0	0.0	0.0	0.0	0.0
Diamond simulant false negative rate (%)	0.0	0.0	0.0	0.0	0.0

Notes: [1] Primary stone set deviates from the standard as a reduced number of stones were analyzed. The Primary sample has a total of 500 mixed stones of which 470 stones were tested due to removal of stones greater than 3.8 mm

- ^[2] Uncertainty is expressed as absolute +/- range and reflects the consistency of the instrument's classification of stones for each of the three trials performed with the ASSURE sample.
- [3] All Synthetic Diamonds reported as referral for this instrument
- [4] Ultra- smalls sample set only tested once



Unimed	CΛ	/ CCEE	ASDI-500
i inimec	NΑ	/ >> FF	A >1 ハー > しい

Date:

May 28th, 2024

Testing ID:

2024-04

INSTRUMENT TESTING SPEED ASSESSMENT

Testing Speed approximates the usage turnaround time that could be expected by a novice user of the diamond verification instrument and is determined by the time required to evaluate the performance of the diamond verification instrument on the Primary Loose stone test set:

- Testing Speed accounts for the time directly associated with stone assessment including loading stones, programming
 any applicable instrument measurement parameters, analyzing the stones, and segregating the analyzed stones into
 respective instrument classified groups.
- Testing Speed does not include the time to initially warm-up the diamond verification instrument (if applicable) nor the time to separate diamonds from synthetic diamonds for each of the instrument classified groups of analyzed stones.
- Testing Speed does not include time associated with interruptions to the testing process.

Diamond verification instruments that continuously load and analyze stones (i.e., autoloading diamond verification instruments) shall also be assessed for a steady-state Instrument Operating Speed.

Testing speed, and instrument operating speed if applicable, are measured in triplicate. The mean value is reported in the Speed Test Results table below. The Uncertainty reflects the absolute +/- range of the results measured over the three trials.

SPEED TEST RESULTS (PRIMARY LOOSE SAMPLE)

Category	Stones per hour	Uncertainty
Testing Speed (all devices)	727	16
Operating Speed (auto-loading devices)	868	8

Notes: none

None

ADDITIONAL FINDINGS

************ End of Report ****************



APPENDIX: Characterization Report for ASDI-500

Date:

May 16th, 2024

Testing ID:

2024-04

Introduction:

The Characterization Report appendix provides additional information about the physical characteristics for the PRIMARY and SMALLS loose sample sets that were either classified by the instrument as Referrals or that were mis-classified (eg. diamond stones incorrectly classified by the instrument as being synthetic diamond). This appendix is only provided to the instrument manufacturer to provide greater insight into instrument performance and will not be published to the ASSURE directory.

All referral stones or mischaracterized stones from the three trials are included in the data tables. If a data table is not included in this appendix, it means that either the instrument performed flawlessly for that category, or that the category was not tested. The stone count figures (shown in brackets) are the total number for the three trials.

Data Table Definitions:

Diamond Referrals: Diamonds that were categorized as Referral.

Non-diamond Referrals: Synthetic Diamonds and Diamond Simulants categorized as Referral.

Diamond False Positive: Synthetic Diamonds and Diamond Simulants incorrectly categorized as Diamond.

Synthetic False Positive: Diamonds and Diamond Simulants incorrectly categorized as Synthetic Diamond.

Simulant False Positive: Diamonds and Synthetic Diamonds incorrectly categorized as Diamond Simulant.

Notes:

Stone counts indicate the number of stones for a category from the three trials (for example, if a specific stone was referred in two of the three trials, it would contribute a stone count of 2 to the characterization table).

Percentage values indicate the corresponding weighted average proportion from the three trials.

Nitrogen Aggregation (% IaB) reflects the proportion of atomic nitrogen present as B-centers (a carbon vacancy surrounded by four substituted nitrogen atoms) for mischaracterized diamonds.



APPENDIX: Characterization Report for ASDI-500

Date:

May 16th,2024

Testing ID:

2024-04

Diamond F	Diamond Referrals - PRIMARY					Fluorescence			
Category		Stone Type		Stren	gth	Colour			
					None (102 of 870)	78.5%			
				Slight (9 of 72)	6.9%	Blue (8 of 48) Red/Org (0 of 0) Yellow (1 of 24)	88.9% 0.0% 11.1%		
		Type I (130 of 1032)	97.0%	Moderate (6 of 42)	4.6%	Blue (6 of 33) Red/Org (0 of 0) Yellow (0 of 9)	100.0% 0.0% 0.0%		
Diamond 100.0	100.0%			Strong (13 of 48)	10.0%	Blue (13 of 48) Red/Org (0 of 0) Yellow (0 of 0)	100.0% 0.0% 0.0%		
(134 of 1050)				None (4 of 18)	100.0%				
					Slight	0.0%			
		Type II (4 of 18)	3.0%	Moderate	0.0%				
				Strong	0.0%				

					Nitrogen Cl	naracterization	
Category		Stone Typ	Stone Type		ion (ppm)	Aggregation (%IaB)	
	100.0%			<100 (23 of 84)	17.7%	0 - 30 (22 of 81) 30 - 70 (0 of 0) >70 (1 of 3)	95.7% 0.0% 4.3%
		Type I	07.00/	100 - 300 (69 of 651)	53.1%	0 - 30 (56 of 612) 30 - 70 (6 of 9) >70 (7 of 30)	81.2% 8.7% 10.1%
Diamond (134 of 1050)		(130 of 1032)	97.0%	300 - 600 (38 of 285)	29.2%	0 - 30 (24 of 165) 30 - 70 (2 of 33) >70 (12 of 87)	63.2% 5.3% 31.6%
				>600 (0 of 12)	0.0%	0 - 30 (0 of 12) 30 - 70 (0 of 0) >70 (0 of 0)	
		Type II (4 of 18)		<100 (4 of 18)	100.0%	0 - 30 (4 of 18) 30 - 70 (0 of 0) >70 (0 of 0)	100.0% 0.0% 0.0%



APPENDIX: Characterization Report for ASDI-500

Date:

May 16th,2024

Testing ID:

2024-04

Diar	nond Ref	errals - SMAL	LS	Fluorescence										
Catego	ory	Stone Type		Streng	Strength		Colour							
				None (150 of 891)	94.9%									
				Slight (1 of 51)	0.6%	Blue (1 of 18) Red/Org (0 of 0) Yellow (0 of 33)	100.0% 0.0% 0.0%							
		Type I (158 of 1032)	100.0%	Moderate (5 of 75)	3.2%	Blue (5 of 51) Red/Org (0 of 0) Yellow (0 of 24)	100.0% 0.0% 0.0%							
Diamond	100.0%			Strong (2 of 15)	1.3%	Blue (2 of 6) Red/Org (0 of 0) Yellow (0 of 9)	100.0% 0.0% 0.0%							
(158 of 1050)		.0%		None (0 of 18)										
											Slight			
		Type II (0 of 18)	0.0%	Moderate										
					Strong									

				N	Nitrogen Characterization			
Category		Stone Type		Concentrati	on (ppm)	Aggregation (%IaB)		
				<100 (29 of 117)	0.0%	0 - 30 (29 of 114) 30 - 70 (0 of 0) >70 (0 of 3)	100.0% 0.0% 0.0%	
	100.0%	Type I		100 - 300 (71 of 633)	30.8%	0 - 30 (63 of 537) 30 - 70 (0 of 0) >70 (8 of 96)	88.7% 0.0% 11.3%	
Diamond (158 of 1050)		,158 of 1032)	100.0%	300 - 600 (58 of 282)	36.7%	0 - 30 (58 of 243) 30 - 70 (0 of 15) >70 (0 of 24)	100.0% 0.0% 0.0%	
				>600 (0 of 0)	0.0%	0 - 30 (0 of 0) 30 - 70 (0 of 0) >70 (0 of 0)		
		Type II (0 of 18)	0.0%	<100 (0 of 18)		0 - 30 (0 of 18) 30 - 70 (0 of 0) >70 (0 of 0)		