



PROJECT ASSURE DIAMOND VERIFICATION INSTRUMENT STANDARD TEST RESULTS

Assessment Report for: Jubilee Diamond Instruments / BELIZE combination tester



Prepared For:

Luc Auer Natural Diamond Council Belgium VOF Hoveniersstraat 22 2018 Antwerpen

Received Date: April 2 , 2024 Assessment Dates: June 4 to June 7, 2024 Testing ID Number: 2024-05 Report Date: July 31, 2024 Approved by:

Quinten Van Avondt Lab Manager

LETTERS & REPORTS: Universiteit Antwerpen (UAntwerp) letters and reports are issued for the exclusive use of the clients to whom they are addressed. No quotations from reports or use of the UAntwerp name is permitted except as expressly authorized in writing. Letters and reports apply only to the specific materials, products or processes tested, examined or surveyed and are not necessarily indicative of the qualities of apparently identical or similar materials, products or processes. The liability of UAntwerp with respect to services rendered shall be limited to the amount of consideration paid for such service and not include any consequential damages. This report or certificate does not relieve sellers/suppliers from their contractual responsibility with regard to the quality/quantity of this delivery, nor does it prejudice clients' right to claim towards sellers/suppliers for compensation for any apparent and/or hidden defects not detected during our random inspection or testing. UAntwerp has not performed a complete analysis of the product. The results contained in this report indicate that the product has passed or failed the specific tests only. These test results, even if rated as "Passed," do not indicate or certify that the product is safe for commercial or consumer use.

ID:

Date: July 31, 2024

DIAMOND VERIFICATION INSTRUMENT

Manufacturer's Name:	Jubilee Diamond Instruments
Instrument Model:	BELIZE Combination Tester (GEM-0022-01)
Serial Number:	S/N: 222404180N
Software Version:	not applicable
Lab Manager:	Quinten Van Avondt
Analyst /Operator:	Cindy De Plukker

Manufacturer stated diamond verification instrument description and features:

- Manual stone feed
- Automatic stone classification (Earth mined diamond, Type IIa/HPHT/CVD diamond, Simulant and Moissanite)
- Manual stone sorting •

Manufacturer stated diamond verification instrument limitations:

- Loose stones •
- All types of mounted jewelry, closed and open-back •
- Stone size 0.02 ct to 10 ct •
- Stone colour D-J
- Stone clarity VVS1 to SI2

INSTRUMENT PERFORMANCE ASSESSMENT

ASSESSMENT CRITERIA

The ASSURE testing methodology and performance metrics are dependent on the operational capabilities of the diamond verification instrument being tested. These are defined by the following three categories:

Category 1- Screen diamonds from synthetic diamonds. This category of device is intended for discrimination of diamonds from synthetic diamonds. It cannot distinguish diamonds from diamond simulants and therefore requires stones to be pre-screened to ensure no simulants are introduced into the device.

Category 2 – Screen diamonds from synthetic diamonds and diamond simulants. This category of device is intended for discrimination of diamonds from synthetic diamonds and diamond simulants. This device cannot distinguish synthetic diamonds from diamond simulants.

Category 3 – Screen diamond from synthetic diamonds from diamond simulants. This category of device is intended for discrimination of diamonds, synthetic diamonds and diamond simulants from each other. This device can distinguish synthetic diamonds from diamond simulants.

Instrument performance for classifying the different kinds of stones was assessed against:

Diamond Verification Instrument Standard Part 1 - Diamond Verification Instrument for Screening Diamonds from

Synthetic Diamonds (2024 03 08)

	Universiteit	APPENDIX: Characterization Report for BELIZE combination tester			
6	Antwerpen			Testing ID:	2024-05

- Diamond Verification Instrument Standard Part 2 Diamond Verification Instrument for Screening Diamonds from Synthetic Diamonds and Diamond Simulants (2024 03 08)
- Diamond Verification Instrument Standard Part 3 Diamond Verification Instrument for Screening Diamonds, Synthetic Diamonds, and Diamond Simulants (2024 03 08)

as referenced in sections 7.3 and 7.4 of the Diamond Verification Instrument Standard – General Requirements for Evaluation Diamond Verification Instruments (2024 03 08). Any deviations from the Standard are noted below:

DEFINITIONS:

Diamond Accuracy Diamond test stones correctly classified as Diamond.	
---	--

	Universiteit	APPENDIX: Characterization Report for BELIZE combination tester			
6	Antwerpen		July 31, 2024	Testing ID:	2024-05

Synthetic Diamond Accuracy	Synthetic Diamond test stones correctly classified as Synthetic				
-,	Diamond.				
Diamond Simulant Accuracy	Diamond Simulant test stones correctly classified as Diamond				
Diamond Simulant Accuracy	Simulant.				
Diamond Referral Rate	Diamond test stones classified as Referral.				
Synthetic Diamond Referral Rate	Synthetic diamond test stones classified as Referral				
Simulant Referral Rate	Diamond simulant test stones classified as Referral				
	Synthetic Diamond / Diamond Simulant test stones incorrectly				
Diamond False Positive Rate	classified as Diamonds.				
Construction Discovery of Fallers Desisting Data	Diamond / Diamond Simulant test stones incorrectly classified as				
Synthetic Diamond False Positive Rate	Synthetic Diamonds				
Discourse de Circulante Falace Davitiva Data	Diamond / Synthetic Diamond test stones incorrectly classified as				
Diamond Simulant False Positive Rate	Diamond Simulants				
Diamond False Nagative Data	Diamonds incorrectly classified as Synthetic Diamonds or Diamond				
Diamond False Negative Rate	Simulants				
Synthetic Diamond False Negative Rate	Fraction of Synthetic Diamonds incorrectly classified as Diamond.				
Diamond Simulant False Negative Rate	Fraction of Diamond Simulants incorrectly classified as Diamond.				
	The average speed at which the diamond verification instrument				
Testing Speed	evaluates the stones in the PRIMARY loose sample set, including set-				
	up time (if any)				
	For auto-loading diamond verification instruments only, the average				
Operating Speed	speed at which stones are evaluated once the instrument achieves a				
	steady-state. Does not include set-up time.				

TEST STONE SETS USED FOR EVALUATION

Loose, Polished Stone Test Sets	Diamond	Synthetic Diamond	Diamond Simulant
Primary Sample Set (>2.0 mm, D-J color)	\boxtimes	\boxtimes	\boxtimes
Supplementary Smalls Sample Set (1.0-2.0 mm, D-J color)			
Mounted, Polished Stone Test Sets	Diamond	Synthetic Diamond	Diamond Simulant
Primary Sample Set (>2.0 mm, D-J color)	\boxtimes	\boxtimes	\boxtimes
Supplementary Smalls Sample Set (1.0-2.0 mm, D-J color)			

Notes: smalls sample not tested

CLEANING PROCEDURE OF STONES PRIOR TO TESTING

Test stones sets are cleaned in an ultrasonic bath of isopropanol for 2 minutes and dried prior to testing to reduce grease and electrostatic charge, as per Section 8 of ASSURE Standard.

ID:

LABORATORY CONDITIONS AT TIME OF ASSESSMENT

Condition	Requirement	Actual
Temperature (°C)	18 to 25°C	22 °C
Humidity (%)	50 to 65%	55 %

RESULTS OF INSTRUMENT PERFORMANCE ASSESSMENT – LOOSE STONES

Performance Metric	Primary	Uncertainty ^[1]
Diamond accuracy (%)	92.8	0.6
Synthetic diamond accuracy (%)	n/a ^[2]	n/a ^[2]
Diamond simulant accuracy (%)	100.0	0.0
Diamond referral rate (%)	7.2	0.6
Synthetic diamond referral rate (%)	100.0 [2]	0.0 [2]
Diamond simulant referral rate (%)	0.0	0.0
Diamond false positive rate (%)	0.0	0.0
Synthetic diamond false positive rate (%)	0.0	0.0
Diamond simulant false positive rate (%)	0.0	0.0
Diamond false negative rate (%)	0.0	0.0
Synthetic diamond false negative rate (%)	0.0	0.0
Diamond simulant false negative rate (%)	0.0	0.0

- Notes: ^[1] Uncertainty is expressed as absolute +/- range and reflects the consistency of the instrument's classification of stones for each of the three trials performed with the ASSURE sample.
 - ^[2] This instrument cannot distinguish between CVD/HPHT and Type IIa diamonds. Stones identified as CVD /HPHT/Type IIa diamonds are classified as REFER.

RESULTS OF INSTRUMENT PERFORMANCE ASSESSMENT – MOUNTED STONES

Performance metric	Primary
Diamond accuracy (%)	85.6
Synthetic diamond accuracy (%)	n/a ^[1]
Diamond simulant accuracy (%)	100.0
Diamond referral rate (%)	14.4
Synthetic diamond referral rate (%)	100.0 ^[1]
Diamond simulant referral rate (%)	0.0

	Universiteit	APPENDIX: Characterization Report for BELIZE combination tester			
6	Antwerpen			Testing ID:	2024-05

Diamond false positive rate (%)	0.0
Synthetic diamond false positive rate (%)	0.0
Diamond simulant false positive rate (%)	0.0
Diamond false negative rate (%)	0.0
Synthetic diamond false negative rate (%)	0.0
Diamond simulant false negative rate (%)	0.0

Notes: ^[1] This instrument cannot distinguish between CVD/HPHT and Type IIa diamonds. Stones identified as CVD / HPHT Type IIa diamonds are classified as REFER.

Two jewelry samples were removed from the sample set as they contained foiled back glass stones which cannot be tested by BELIZE.

INSTRUMENT TESTING SPEED ASSESSMENT

Testing Speed approximates the usage turnaround time that could be expected by a novice user of the diamond verification instrument and is determined by the time required to evaluate the performance of the diamond verification instrument on the Primary Loose stone test set:

- Testing Speed accounts for the time directly associated with stone assessment including loading stones, programming any applicable instrument measurement parameters, analyzing the stones, and segregating the analyzed stones into respective instrument classified groups.
- Testing Speed does not include the time to initially warm-up the diamond verification instrument (if applicable) nor the time to separate diamonds from synthetic diamonds for each of the instrument classified groups of analyzed stones.
- Testing Speed does not include time associated with interruptions to the testing process.

Diamond verification instruments that continuously load and analyze stones (i.e., autoloading diamond verification instruments) shall also be assessed for a steady-state Instrument Operating Speed.

Testing speed, and instrument operating speed if applicable, are measured in triplicate. The mean value is reported in the Speed Test Results table below. The Uncertainty reflects the absolute +/- range of the results measured over the three trials.

	Universiteit	APPENDIX: Characterization Report for BELIZE combination tester			
6	Antwerpen			Testing ID:	2024-05

Category	Stones per hour	Uncertainty
Testing Speed (all devices)	256	12
Operating Speed (auto-loading devices)	n/a ^[1]	n/a ^[1]

Notes: ${}^{\scriptscriptstyle [1]}{} \, \text{not}$ applicable for this device, the device has manual feed

ADDITIONAL FINDINGS

None